

VEER SURENDRA SAI UNIVERSITY OF TECHNOLOGY, ODISHA

P.O: Engineering College Burla (Siddhi Vihar), Dist: Sambalpur
Odisha-768018, India

Website: www.vssut.ac.in, E-mail: registrar@vssut.ac.in, Ph:(0663)2430573,Fax-2430592



No. VSSUT/ CHEM/

/2026

Date: 29 / 09 /2026

QUOTATION CALL NOTICE

Sealed quotations are invited from the intending reputed wholesaler/Direct Importers /Registered Authorized Dealers with GST Registration certificate, copy of PAN card for supply of **Instruments** for **Chemistry Department of VSSUT, Burla**. This would reach the office of the undersigned through speed or registered post by **31. 10. 2026 (5 PM)**. Detailed information and the tender document are available in the University website (www.vssut.ac.in) or contact PI (9937082348), DST-FIST Project.


REGISTRAR

Memo No. VSSUT/CHEM/

/2026

Dated 29 / 9 /2026

Copy to:

1. The Director, I & PR, Govt. of Odish with a request to publish the above advertisement in one issue of the Odisha daily edition (ALL ODISHA DAILY EDITION) of "THE SAMAJA" and THE ALL-INDIA EDITION of "THE TIMES OF INDIA" a using minimum space at I & PR approved/lowest rates on or before **01.10.2026**. The bills may be sent to the Comptroller of Finance, VSSUT, Burla- 768018 in TRIPLICATE along with copies of the paper in which the publication is made, for necessary payment.
2. Dean F&P, with a request to hoist the tender call notice in the University web site www.vssut.ac.in for wide publicity.
3. Steno to Registrar for record.
4. COF, for information and record.
5. Dean SRIC, for information and record.
6. HOD, Chemistry for information and record.
7. Prof. S K Swain PI, DST-FIST Project for information and record.
8. Notice Board, VSSUT Burla for widely circulation.


REGISTRAR
REGISTRAR
V.S.S. University of Technology,
Odisha, Burla





SECTION I: INVITATION FOR BIDS (IFB)

Sealed Bids (properly stitched separately) in two separate covers (Technical Bid and Price Bid) are invited by the **"The Registrar, Veer Surendra Sai University of Technology, Odisha"** from the manufacturers/authorized distributors/ dealers for supply of Instrument, equipment, machineries etc. of reputed make (National/ International) for the Department of Chemistry, VSSUT Burla, Odisha.

The Bidders may download the Tender Documents directly from the website available at <http://www.vssut.ac.in> and the Tender cost fee of Rs.1000/ (Nonrefundable) by way of separate Demand Draft drawn in favor of **"The Comptroller of Finance, Veer Surendra Sai University of Technology, Burla"** payable at SBI, Burla should be enclosed along with the Bid. The Tender cost fee and the EMD amount of 3% of the quoted amount or **RS: 3.0 Lakh (Refundable)** whichever is less should be paid in favor of **Comptroller of Finance, Veer Surendra Sai University of Technology, Burla Sambalpur** should be submitted separately in **separate demand drafts**. In case of any bid clarification, responsibility lies with the bidders to collect the same from the website and the purchaser shall have no responsibility for any delay/ omission on part of the bidder.

TIME SCHEDULE:

- a) Price of bidding document: **Rs. 1000/- (Non-refundable).**
- b) Date of commencement of downloading bidding document **01. 10. 2026 at 4 PM.**
- c) Last date and time for Receipt of bids **31. 10. 2026 up to 5 PM.**
- d) PLACE OF OPENING OF TENDER AND ADDRESS FOR COMMUNICATION AND RECEIPT OF BID DOCUMENTS:

THE REGISTRAR
VEER SURENDRA SAI UNIVERSITY OF TECHNOLOGY, ODISHA
At- Burla, Po-Burla Engineering College, Dist-Sambalpur-768018,
Tel. No-0663-2430211 Fax No-0663-2430204

Sd/-
REGISTRAR
VSSUT Burla
REGISTRAR
V.S.S. University of Technology,
Odisha, Burla

SECTION-II: GENERAL TERMS AND CONDITIONS

GENERAL TERMS & CONDITIONS OF CONTRACT FOR SUPPLY, INSTALLATION AND DEMONSTRATION OF THE INSTRUMENTS, EQUIPMENT REQUIRED FOR DEPARTMENT OF CHEMISTRY, VSSUT, BURLA, ODISHA.

1. Document Establishing Bidder's Eligibility & Qualification

The Bidders shall furnish as part of the Bid the following Documents establishing Bidder's eligibility and qualification to the Purchaser's satisfaction. Non-compliance of any of these conditions shall disqualify the eligibility.

- 1.1 Manufacturer / Authorized Distributor / Dealer having valid license / certificates for the quoted item and the direct Importers holding valid Import License Manufacturer / Authorized Distributor / Dealer of the product are eligible to participate in the Bid.
- 1.2 The Bidder whether manufacturer/ distributor/ dealer must have experience of supply and installation of the quoted items in **different IITs/ NITs/ Institutes of national repute** during last preceding 3(Three) years reckoned from the date of bid opening and the details must be submitted along with documentary proof.
- 1.3 The Bidders shall have to produce document in support of their service associates **nearest to Bhubaneswar/ Sambalpur**, Odisha.
- 1.4 Bidder shall have to provide operational Training for **Machineries** to one Official of each consignee at Suppliers Plant or Consignee site **at least for 7 days whichever is suitable to Purchaser**.
- 1.5 The Bidder shall quote item of one reputed Brand/model with all accessories in complete to perform functionality of Equipment/Machinery.

2. Document Establishing Goods Eligibility

The instruments and equipment offered against the schedule of requirement of instruments, equipment and Machineries should be in accordance with the stipulated specifications and of one reputed brand/model (N.B: **Variation in specification is allowed up to $\pm 1\%$ in case of Equipment**)

- 2.1 The documentary evidence establishing the brand and the model may be in the form of literature, pamphlets, manuals, drawing, circuit diagram etc.
- 2.2 Detailed description of instruments and equipment with essential technical and performance characteristics may also be furnished.



3. Technical Bid (COVER - A)

The following document should be submitted in cover-A.

- 3.1 Earnest Money Deposit.
- 3.2 Technical details of the equipment as per Annexure-V.
- 3.3 Copy of the manufacturing license/ import license/ Authorized Distributor/ Dealer certificates
- 3.4 Copy of the authorization from the Manufacturing Company in case of Authorized Distributor /Dealer.in Annexure-III along with Manufacturer Industry Registration and Tax Registration Certificate.
- 3.5 VAT/ST clearance certificate up to **date** where applicable.
- 3.6 Copy of the IT PAN Card.
- 3.7 Detail name, address, telephone no. fax, and e-mail of the firm and of the Director / Managing Director / Proprietor of the firm (As per Annexure IV)
- 3.8 Address, Telephone No., e-mail, Fax of the Branch Office/ Contact Person/ Liaisoning Office in Odisha.
- 3.9 Power of Attorney/ Authorization to a person for liaisoning and monitoring the business on behalf of the manufacturer / bidder but not entitled to raise the bills.
- 3.10 **The original bid document signed & sealed by authorized person in each page as a token of acceptance of all terms and conditions of the tender with original receipt.**
- 3.11 Any deviation in the specification of the item including standard accessories / optional accessories in complete for functionality of Machine should be marked in **bold letters**. (**N.B: Variation in specification is allowed upto±5% in case of Machineries/Equipments**)

4. Price Bid (COVER – B)

- 4.1 The hard copy of price bid giving the rates for various instruments & equipment and other items should be submitted along with sealed soft copy of **price bid in Excel format through CD/Pen drive** both in separate sealed cover here in after called **Cover B (Price Bid)**. **Price Bid (Cover - B)** of the bidders who qualify in **Technical Bid (Cover – A)** will only be opened and will be communicated through **E-mail/Fax**.
- 4.2 The price of the each item shall be quoted as per the prescribed Price Schedule Format at Annexure-I (a) and I (b) along with price break up of custom duty, Excise Duty, CST, Packing, Forwarding and Handling charges, Insurance charges, ET, Freight up to destination including unloading, VAT, commissioning including testing and training with total price per item at **FOR** destination. The bidders are required to submit the individual price of each instrument(s) and equipment(s) as indicated in the schedule of requirements.



REGISTRAR
V.S.S. University of Technology.
Odisha, Burla

4.3 The **Cover B** of the technically qualifying bidders shall be only opened at the Office of the “**The Registrar, Veer Surendra Sai University of Technology, Burla**” on the date and time to be communicated to them after technical evaluation of **Cover A** by E-mail/Fax..

5 BID CONDITIONS

- 5.1 The bidders should verify the sites of existing laboratories of Department of Chemistry, VSSUT, Burla and the proposed lay out Plan indicating the location of each unit for necessary Technical Evaluation. The scope of Supply as mentioned in the schedule of requirements if not sufficient for full function of the Equipment/Machinery should be intimated in writing with the technical bid.
- 5.2 The quoted rate shall not vary with the quantum of order placed or destination point.
- 5.3 A copy of the original bid conditions and the schedules should be signed by the bidder at the bottom of each page with the office seal duly affixed and returned along with the bid. Bid schedule should be duly filled in with an **index** and **page number** for the documents, enclosures & EMD etc. **Paging** must be done for all the documents submitted.
- 5.4 Bids should be type written or Computerized and every correction/ over writing in the bid should invariably be attested with signature of the bidder with date before submission of the bids to the authorities concerned. No revision of price upward or downward will be allowed once the bid is opened. However the purchaser shall have the right for considering the exchange rate of foreign currencies on verification of documents.

5.5 Language of Bid

The Bid prepared by the bidders and all correspondence and document relating to the bid exchanged by the Bidders and the *Purchaser*, shall be written in the English language. Supporting document and printed literature furnished by the Bidders may be written in another language provided they are accompanied by an accurate translation of the relevant passages in the English language in which case, for purposes of interpretation of the Bid, the English translation shall govern.

5.6 Bid Price

- The contract shall be for the full quantity as described above. Corrections, if any, shall be made by crossing out, initialling, dating and re-writing.
- All duties, taxes, and other levies payable on the raw materials and components, job contract shall be included in the total price.
- VAT in connection with the sale shall be shown separately.




REGISTRAR
V.S.S. University of Technology,
Odisha, Burla

- The rates quoted by the bidders shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- The price shall be quoted in Indian Rupees only.

5.7 GST clearance

Copies of valid GST clearance Certificates shall be furnished by the Bidders and the originals of the above certificates shall be produced to the purchaser before placement of notification of award if asked for by the Purchaser.

5.8 EMD

All bidders are required to submit EMD **not less than 3% of the quoted amount or 3.0 Lakhs whichever is less** in shape of **Demand draft** drawn in favour of **"The Comptroller of Finance, Veer Surendra Sai University of Technology, Burla"** payable at Burla only. The EMD shall be in Indian Rupees.

NOTE: Non-submission of EMD or submission of less EMD than the desired one shall result in rejection of Bid. The EMD deposited against other Bids cannot be adjusted or considered for this Bid. No interest is payable on EMD.

5.9 SUBMISSION OF BIDS

Sealing and Marking of Bids

Bid should be submitted in two Bid system containing two parts as detailed below.

Sealed Cover-A: Technical Bid.

Sealed Cover-B: Price Bid (hardcopy & sealed soft copy in CD/pen drive)

Both the sealed envelopes should then be put in one outer cover and each cover should have the following indication:

- Reference No of Bid _____
- Bid regarding _____
- Due date & time for submission of the Bid _____
- Due date & time for opening of the Bid _____
- Name of the Firm _____

NOTE:

A. Bids submitted without following two Bid system procedures as mentioned above will be summarily rejected.

B. Please Note that prices should not be indicated in the Technical Bid. The Prequalification document including EMD as required in the Bid document should invariably be accompanied with the Technical Bid (Cover A).

The outer envelope shall indicate the name and address of the bidders to enable the bid to be returned unopened in case it is declared "late". If the cover containing the outer envelope is not sealed and marked as required, **Purchaser** will assume no responsibility for the bid's misplacement or premature opening.

The above procedure shall be adopted both for the Technical bid and price bid separately. Telex, cable, email or facsimile bids will be rejected.

5.10 Deadline for Submission of Bids

Bids must be received by the **Purchaser** at the address specified not later than the time and date specified in the Invitation of Bids. In the event of the specified date for the submission of bids being declared a holiday for the **Purchaser**, the bids will be received up to the appointed time on the next working day.

The **Purchaser** may, at its discretion, extend this deadline for submission of bids by amending the bid document, in which case all previous rights and obligations of the purchasers and bidders will remain same till the extended date.

5.11 Modification and Withdrawal of Bids

No Modification and Withdrawal of Bids is allowed between the interval of time of submission and the last date and time of the bids.

No bid may be withdrawn in the interval between the deadline for submission of bids and the expiration of the period of bid validity specified by the bidders on the bid form.

5.12 BID OPENING

5.13 The **Purchaser** will open technical bids, in the presence of bidder's representatives who choose to attend at the Office of the "**The Registrar, Veer Surendra Sai University of Technology, Burla**".

5.14 The bidder's representatives who are present shall sign a register evidencing their attendance. In the event of the specified date of bid opening being declared a holiday for the **Purchaser**, the bids shall be opened at the appointed time and location on the next working day.

5.15 The bidder's names, and the presence or absence of the requisite EMD and such other details as the **Purchaser**, at its discretion, may consider appropriate will be announced at the opening. No bid shall be rejected at bid opening, except for late bids, which shall be returned unopened to the bidders.



7


REGISTRAR

V.S.S. University of Technology,
Odisha, Burla

5.16 Acceptance of the Bid

- Bidders submitting bids would be considered who have considered and accepted all terms and conditions. No enquiries, verbal or written, shall be entertained in respect of acceptance or rejection of the bid.
- Genuine equipment and instrument etc. should be supplied. Bidders should indicate the source of supply i.e. name and address of the manufacturers from whom the items are to be sourced.
- Supply of equipment means – Installation and Commissioning (except civil works), Demonstration as well as Training at site. **No separate charges will be paid on this account.**

5.17 Rejection of the Bid

The Bid document shall be out-rightly rejected under following stipulation and no correspondence will be entertained whatsoever.

- If the Bidders has not furnished the required **Tender paper cost** and **EMD** or EMD exemption certificate from competent authority.
- If the Bidders has not submitted the Price as per the prescribed format **Annexure-I(a) and I (b)**
- **Manufacturing Authorization Annexure-III and in case of Authorized Distributor / Dealer/Suppliers must submit Manufacture authorization along with copy of above documents of Manufacture Industry.**
- If the bid is not supplemented with breakup of standard accessories / Optional accessories & cost of AMC separately for three years after completion of warranty period **(In case of major machinery only).**
- Photo copy of the up-to-date valid manufacturing license/ import license (if it is imported) /dealership certificate/Distributor certificate of the product along with Tax registration Certificate of Manufacturer issued from competent authority.
- If the bidders, whether manufacturer or authorized distributor/ dealer have not supplied the required quantity for qualification as per the eligibility criteria and not submitted the performance statement at **Annexure-IV with supporting documents.**
- If the quoted product of the bidders do not confirm to technical specification with complete accessories for functional Equipment/Machinery and standard of workmanship required by the Purchaser.
- If the bidder has not furnished technical details of the equipments and machinery with **one make & model** as per **Annexure-V.**
- **If bidder will quote items of more than one make/model.**

- If the bidder has not furnished detailed mandatory drawings, **catalogue/**, Foundation drawings & schedule of supply of items, if required.
- If the bidders have not agreed to give **bid validity**.
- If Bidder is not willing to provide operational Training for Machineries to one Official of each consignee at Suppliers Plant or Consignee site **at least for 7 days in case of major machineries whichever is suitable to Purchaser.**

5.18 Purchaser's Right to accept any Bid and to Reject any Bid

The Purchaser reserves the right to accept or reject any bid and to annul the bidding process and reject all the bids without assigning any reason thereof at any time prior to award of Contract, without thereby incurring any liability to the affected Bidders or Bidders on the grounds of such action of the purchaser. In case no bidder qualifies as per qualifying criteria and standards, purchaser may at his discretion relax qualification criteria for award of contract.

5.19 Evaluation and Comparison of Bids

The comparison shall be of FOR destination price basis including the price of all costs wherever applicable as well as duties and taxes (**but excluding VAT**) paid or payable on Machineries, instruments & equipment incorporated or to be incorporated in the items including the warrantee/guarantee period from the date of installation.

- The Purchaser's evaluation of a bid will take into account, in addition to the bid price and the price of incidental services.
- The purpose of bid evaluation is to determine substantially responsive bid with the lowest evaluated cost, but not necessarily the lowest submitted price, which should be recommended for award.
- Evaluation of bids should be made strictly in terms of the provisions in the bid document to ensure compliance with the commercial and technical aspects.
- The past performance of the suppliers will be taken into account while evaluating the bids.
- Cost of the inland transportation, insurance and other costs within the Purchaser's Country incidental to delivery of the goods to their final destination;
- **Alternative options of offer shall not be allowed.**
- Each Bidder shall submit only one quotation with one make & model.
- The quotation would be evaluated separately for each item
- Sales Tax in connection with sale of goods shall not be taken into account in evaluation.

- Willingness to provide operational Training to one Official of each consignee at consignee site for at least **7 days whichever is suitable to Purchaser.**

6.0 Supply Conditions

6.1 Delivery of Goods

The delivery of goods shall be made by the supplier to the Consignee in accordance with the order placed as shall be detailed in the Schedule of requirements & technical specifications.

6.2 Inspection/ Test/Training

The inspection/examination/ test may be conducted in the premises of the Supplier or at the goods final destination or at the premises of the consignee, as will be decided by the Purchaser.

The purchaser's right to inspect/ examine/test & where necessary to reject the instruments & equipment after the arrival of the goods at the final destination, shall in no way be limited or waived by the reason of the goods having been inspected and tested by the manufacturer previously. In case of rejection of the goods at the final destination after inspection and test as stipulated above and in case any inspected/ tested goods fail to confirm to the specification/ working condition, the purchaser may reject them and the supplier shall replace/ repair the same free of cost.

6.3 Warrantee Period (comprehensive)

The Bidders must quote the period of comprehensive **warranty** from the date of completion of the satisfactory commissioning as per (**Annexure-II**). This also includes all accessories related to instruments & equipment quoted for.

6.4 Up time Warrantee

The Bidders should provide uptime guarantee of 95%.

6.5 Downtime Penalty Clause

During the warranty period, desired uptime of 95% of 365 days (24 hours) if downtime exceeds 5%, penalty in the form of extended warranty, double the number of days or more will be applied for which the equipment goes out of service.

In no case the machineries should remain in non-working condition for more than 30 days beyond which a penalty of 0.2% of machine cost will be charged per day.

6.6 Payment Terms

No advance payment will be made by the Purchaser to the supplier for performance of the contract. 100% of the contract price shall be paid within 30 (thirty) days after satisfactory supply, installation, demonstration, Commissioning & training and stock entry of bills of the goods within due date of delivery.



6.7 Transportation

The Supplier shall be required to meet all transport and storage expenses until commissioning of the instrument(s) / equipment covered in the contract.

6.8 Taxes and Duties

The Supplier shall be entirely responsible for payment of all Taxes, Duties etc. incurred until delivery of the contract goods to the Consignee subject to recovery afterwards in the bill as claimed in the Bid offer.

VAT as applicable is payable, to the suppliers of the State of Odisha if claimed in the Bid offer.

VAT/ CST will be paid to the Suppliers of the outside State other than Odisha, if claimed in the Bid offer.

Any revision of VAT/ CST shall automatically be taken into account.

Entry Tax, if paid by the Supplier, at the local (destination head) Corporation/ Municipality/ NAC is allowed once only on production of money receipt for such payment, if claimed in the Bid offer. Any other statutory levy imposed by the Govt. of India/ Govt. of Odisha from time to time will be considered extra on demand with adequate proof thereof. The service tax and the work contract tax shall be levied (Wherever applicable). Income Tax as applicable shall be deducted at source.

6.10 Period of Validity of Bids

- The bid rates should be kept open/ valid for a period of **180** days from the date the Bids are opened.
- A bid valid for a shorter period i.e less than **180** days shall be rejected, as nonresponsive.
- In absence of any indication of the date of validity in the bid, it will be presumed that the offer will remain valid for the minimum period i.e. **180** days as prescribed above.

6.11 Commissioning Period

Maximum commissioning period is **30 days** from the date of supply OR **120 days from the date of issue of Purchase Order** failing which the purchaser will have the right to impose penalty for the delay period @ **0.5% per week of the contract value of item/items excluding taxes from the bill amount subject to maximum of 10%**. However Registrar has right to extend the delivery period/commissioning period in special cases.

6.12 Penalty against Non Supply

In case of non-supply of Stores within the due date i.e. within the date of delivery the EMD deposited by the bidder shall be forfeited.

6.13 Jurisdiction of the Court



11

REGISTRAR

**V.S.S. University of Technology,
Odisha, Burla**

The Purchaser and the Supplier shall agree that the competent Court at Sambalpur shall have the jurisdiction to try and decide anything between the parties and they may approach the Competent Court at Sambalpur if required at any time.

**REGISTRAR
VSSUT, Burla**


**REGISTRAR
V.S.S. University of Technology,
Odisha, Burla**

12



[Faint, illegible text]

SECTION – III: FORMS AND ANNEXURES

ANNEXURE-I (a)

PRICE SCHEDULE (ITEM WISE) FOR EQUIPMENT

Item Serial No	Item Description	Country of origin with make & model	Quantity & Unit							
				Exwarehouse/ Ex-showroom/ Off-the-shelf (a)	Packing & Forwarding (b)	Inland transportation, insurance and other local costs incidental to delivery (c)	GST as applicable (d)	Unit price (a+b+c+d)	Total Price (a+b+c+d)	

ANNEXURE-I (b)

PRICE SCHEDULE (ITEM WISE) FOR EQUIPMENT

Item Serial No	Item Description	Country of origin with make & model	Quantity & Unit							
				Exwarehouse/ Ex-showroom/ Off-the-shelf (a)	Packing & Forwarding (b)	Inland transportation, insurance and other local costs incidental to delivery (c)	GST as applicable (d)	Unit price (a+b+c+d)	Total Price (a+b+c+d)	




REGISTRAR
V.S.S. University of Technology,
Odisha, Bhubaneswar

ANNEXURE - II

WARRANTY MAINTENANCE CONTRACT AGREEMENT.

THIS AGREEMENT made the.....day of, 20__ between the **"The Registrar, Veer Surendra Sai University of Technology, Burla"** (hereinafter "the Purchaser") of the one part and M/s..... (here in after called "the Supplier") of the other part:

WHEREAS the Purchaser invited bids for certain Goods & ancillary services viz, supply and commissioning of the instruments & equipment at Bhubaneswar including Comprehensive Warranty Maintenance Services and has accepted a bid by the Supplier for the instruments & equipment specified below at the Consignee site including Comprehensive Warranty maintenance Services for a period of 2 (Two) year from the date of installation & commissioning of the instruments & equipment as per award of Contract No..... dated

Name of the Equipment & machineries Qty

(To be filled in as per details of goods in the award of Contract)

NOW THIS AGREEMENT WITNESSETH AS FOLLOWS:

1. Maintenance Services shall consist of Preventive and Corrective maintenance of equipment specified above & will include repair and replacement of parts free of cost.
2. Preventive maintenance, monthly once, which includes:
 - 2.1 Check-up to ensure that device connection is proper, cabling is at proper condition etc.
 - 2.2 Cleaning of the above instruments & equipments and checking the System Performance.
3. The Supplier is to furnish the tentative schedule of the preventive maintenance of Warranty Maintenance Contract (WMC) to be carried out.
- 4 The parts replaced must be new parts or equivalent in performance to new parts.
5. The Supplier will also provide the same maintenance service in case of the movement of equipment from the place of original installation to a different place or location, if the equipment is shifted by the Purchaser to another place or location at the cost and risk of the purchaser.
6. Any complaint informed through telephone must be acknowledged with a Complaint No. by the Supplier which will be noted by Consignee. All further contact with the Supplier on such complaint will be initiated through that Complaint No. Once rectification done, that No. will be cancelled by both parties. A register is to be maintained by the Supplier where complaints are to be noted along with Complaint No.
7. The maintenance shall normally be done at the earliest.
8. The Service Engineer of the Supplier will be allowed to handle the respective plant & machineries only in presence of the officer in charge at the Consignee site.
9. The Supplier should ensure that maintenance job is not hampered/delayed due to paucity of spares/inadequate manpower etc.




REGISTRAR
V.S.S. University of Technology.
Odisha, Burla

10. The Supplier should submit the services call report, to the Consignee for each and every service call without fail.

11. The Supplier evaluation data format for the WMC of Consignee systems may be filled up for necessary action.

12. All formats after filled up should be signed at the end of each page by the Supplier.

13. After completion of the work/repair/maintenance, the Purchaser shall issue a certificate of completion to the supplier to that effect.

Signature

For the Purchaser

Name:

Designation:

Address:

Telephone No:

Signature

For the Supplier

Name:

Designation:

Address:

Telephone No:



REGISTRAR
19.S.S. University of Technology,
Odisha, Burla

ANNEXURE-III

MANUFACTURES' AUTHORISATION FORM

No. _____ / Date _____

To

The Registrar,
VSSUT Odisha
Burla, Sambalpur.

Dear Sir, Bid No. _____

We _____ who are established and reputable manufacturers
of _____ having factories at _____ (Address of Factory) do thereby authorize
M/s. _____ (Name and address of Agent) to submit a bid and sign the
contract with you against the above bid.

* No company or firm or individual other than M/s. _____ are authorized to bid and
conclude the contract in regard to this business against this specific invitation for bid.

We hereby extend our full guaranty and warranty as per general conditions of contract for the goods and
services offered by the above firm against this bid.

Yours faithfully,
(Signature for and on behalf of Manufacturers)

Note: This letter of authority should be on the letterhead of the manufacturer and should be signed by a person,
competent and having the power of attorney to bind the manufacturer. It should be included original by the
Bidders in its bid.

- This para should be deleted for simple items where manufacturers sell the product through different
stockiest.

16

REGISTRAR

V.S.S. University of Technology,
Odisha, Burla

ANNEXURE-IV

DETAILS OF THE BIDDERS

Bid Reference No.

Name and address of the Bidder:

- 01 Name of the bidder
 - a) Full postal address
 - b) Full address of the premises
 - c) Telegraphic address
 - d) Telephone number
 - e) Fax number
 - f) E mail:
 - g) PAN No
 - h) TIN No
- 02 Total annual turn-over (value in Rupees)
- 03 Quality control arrangement details
- 04 Test certificate held
 - a) Type test
 - b) BIS/ISO certification
 - c) Any other
- 05 Details of staff
 - a) Technical
 - b) Skilled
 - c) Unskilled
- 06 Branch Office/ Contact Person/ Liaisoning Office in Odisha.
 - a) Address
 - b) Telephone No.
 - c) e-mail,
 - d) Fax

Signature and seal of the Bidder



17



REGISTRAR
V.S.S. University of Technology,
Odisha, Burla

ANNEXURE-V

Technical details of the Equipments to be supplied by the bidder

Bid Sl No. of the item	Tender specification	Bidders Specification with make and model no (Enclose manufactures catalogue / brochure for each item)	Deviation if any With university specification

Signature and seal of the Bidder


REGISTRAR
V.S.S. University of Technology.
Odisha, Burla

18

KASHI K. J. J. J.
Registrar, V.S.S. University of Technology,
Odisha, Burla



SECTION-IV: SCHEDULE OF REQUIREMENTS AND TECHNICAL SPECIFICATIONS

Sl. No.	Name of instruments/ Equipment's/ Quantity	Specification
01	Generalized Specifications for Fluorescence Spectrophotometer Quantity (01)	Price shall be inclusive of packing, forwarding, freight, delivery at door, installation and demonstration with warranty inclusive of all duties and taxes. Light source : 150 W Xenon lamp (Light source lifetime : 2000 hours) Monochromator : Blazed holographic grating, 1300 grooves/mm Wavelength scale : 200 to 900 nm and 0 order Scanning Wavelength range : 200 to 900 nm and 0 order Spectral bandwidth : Excitation: 1.5, 3, 5, 10, 15, 20 nm Emission: 1, 3, 5, 10, 15, 20 nm Resolution: 1.0 nm or less (Emission) Wavelength accuracy : Within ± 1.0 nm Detector : PMT offers scan wavelength range to 900nm Sensitivity: The SN ratio of the Raman line of distilled water is 350 and more (P-P); 1,000 and more (RMS); (Excitation 350 nm, Spectral bandwidth 5 nm, Response 2 sec) Wavelength scanning: Max. 60,000 nm/min Wavelength slewing speed: 60,000 nm/min. Interface : USB 2.0/3.0 The user interface is easy to understand and use. - The instrument should be supplied with software that has been tested and confirmed to run on 64-bit versions of the Windows 11 Professional operating system is provided with the instrument. Functions for converting to CSV files/text files are also provided. - The standard features provided by the software include a spectrum mode, 3D spectrum mode, photometric mode, quantitation mode, and time-course mode. Validation software is also included. - Enhanced sensitivity and dynamic range enable fluorescence as well as bioluminescence, chemiluminescence, and electro-luminescence measurements. - High-speed 3D scanning enables rapid acquisition of 3D spectra. - Spectral-Corrected Excitation and Emission spectra can be scanned. - Fluorescence quantum yield and Fluorescence quantum efficiency measurements are possible with future attachment of Integrating Sphere Unit. - One no. Non Fluorescent cell must be supplied along with the instrument



		• Branded i3 PC with Original Windows 11 Professional must be supplied alongwith the system • 1.5 KVA online UPS with more than 45 minutes backup. • Vibrational- fewer tables 5*3 ft. • Must be supplied two pair of fluorescence cuvette. • Warranty: 2 years from the date of installation • Installation & training: is to be provided by the vendor free of cost at site. Price shall be inclusive of packing, forwarding, freight, delivery at door, installation and demonstration with warranty inclusive of all duties and taxes.
02	Generalized Specifications for Thermo Gravimetric Analyzer TG-DSC Quantity (01)	STA TECHNICAL SPECIFICATIONS • Simultaneous Thermo Gravimetric Analyzer TG-DSC • DSC mode for quantitative studies • Horizontal differential balance design • Differential Thermocouple Method • Built in mass flow controller for two gases • Typical Sample amount max 200mg • Auto sampler Max 50 Samples (Optional) • High temperature furnace upgradeable upto 1500 deg C • Sample observation Real View (Patented) upgradeable • Furnace cool down time: Less than 20 minutes • Temperature range: RT to 1100 degrees C • Programmed scanning rate : 0.01 degrees to 150 degrees C/min • Temperature accuracy: ± 0.2 degrees C • Temperature precision: ± 0.07 degrees C • TG Balance dynamic range: ± 400 mg • TG Sensitivity: $0.2 \mu\text{g}$ • TG Resolution: $0.05 \mu\text{g}$ • TG Accuracy: $\pm 0.1\%$ • TG Precision: $\pm 0.04\%$ • TG Drift: $< 10 \mu\text{g}$ • TG Stability: $< 10 \mu\text{g}$ • TG Repeatability: $< 10 \mu\text{g}$ • TG Environmental Temperature Stability: $< 3 \mu\text{g/ degrees C}$ • DTA measurement range: $\pm 1000 \mu\text{V}$ • DTA Sensitivity: $0.06 \mu\text{V}$ • DSC Calorimetric Precision: $\pm 3\%$ or better • Temperature Modulation: Total heat flow specific heat component and kinetic component • Gas flow rate: 0 to 1000 ml/ min with MSF (Mass flow controller) • Nitrogen cylinder with low pressure regulator • Software: 1. Kinetics Analysis (STA and DSC)

		2. Peak Separation (STA and DSC) 3. Profile Separation (STA and TMA) 4. Heating Rate Simulation (STA and DSC) 5. Auto Analysis 6. Specific Heat Calculation (STA and DSC) 7. Temperature Modulated Analysis (STA and DSC) 8. Warranty: two years 9. Branded i3 PC with Original Windows 11 Professional must be supplied alongwith the system 10. 1.5 KVA online UPS with more than 45 minutes backup. 11. Vibrational- fewer tables 5*3 ft. Price shall be inclusive of packing, forwarding, freight, delivery at door, installation and demonstration with warranty inclusive of all duties and taxes
03	Generalized Specifications for Electrochemical Workstation Quantity (01)	1. Modular, Independently Addressable 2-Channel Electrochemical Workstation 2. Each channel must operate independently and simultaneously 3. Compliance voltage: ± 12 V or better 4. Maximum applied voltage: ± 10 V or better 5. Maximum output current: ± 500mA 6. Current ranges: 10nA to 1A (minimum 8 ranges) with automatic current range selection. 7. Current resolution: 800fA 8. Future possibility to upgrade the same unit to low current up to 1pA & also for high current of 25A or better 9. Voltage accuracy applied and measured: $< \pm 1$ mV ± 0.03 % of setting 10. Frequency range from 10 μHz to 11 MHz (for at least 1 channel) 11. Future expandable facility of high-frequency capability up to 11Mhz or more 12. Accuracy of 1% and 1° at 1 Hz with contour plot support. 13. Accuracy of applied and measured current: $< 0.1\%$ FSR. 14. Potentiostat rise/fall time: < 500 ns 15. Resolution of applied potential: 5 μV or better 16. Measured current resolution: 0.004% of the range or better 17. Measured Voltage Resolution: 1μV at 60mV range or better 18. Input Impedance :1TΩ 19. Acquisition time: 12 μs or better 20. Input Bias Current: < 5 pA or better 21. Maximum scan rate : 200 V/s or better. 22. Electrode connections: 2,3,4,5 23. Future upgradability for RRDE application with same unit.




24. Real-time update of experimental parameters without pausing/stopping the experiment.
25. **Modify experimental protocols 'on the fly'.**
26. Capability to control and connect external devices.
27. **Second voltage measurement – can measure the voltage of the negative electrode simultaneously.**
28. Ewe vs and Reference and Ece Vs Reference can be measured simultaneously
29. High Precision Coulombic Efficiency Determination up to 10 ppm or less
30. Internal Resistance Determination
31. Both Ground and floating mode should be available

Electrochemical Impedance Spectroscopy (EIS)

1. Real-time fit and simulation with Nyquist, Bode, Admittance, Dielectric, and Mott-Schottky plots.
2. EIS quality indicators for measurement validation (**THD, NSD, NSR**) — enabling validated high-fidelity impedance measurements across a wide frequency range.
3. **Patented and unique transient state correction technique - Drift correction** for long-duration battery or corrosion studies.
4. Sensitivity functions for accurate circuit element fitting.

Software Requirements

1. Pre-installed with ≥80 electrochemical techniques.
2. Support for electrochemistry techniques including Cyclic Voltammetry (CV), Linear Sweep Voltammetry (LSV), Electrochemical Impedance Spectroscopy (EIS), Chronoamperometry, and Chronopotentiometry.
3. Support for GITT and PITT methods.
4. Galvanostatic charge/discharge including C-rate control with various variable representations.
5. Industrial CC-CV (Constant Current – Constant Voltage) operation.
6. Current scan, voltage scan, constant power/resistance modes.
7. Multigraph and customizable variable plots for each axis.
8. At least 3 limits and 3 recording conditions per sequence or cycle (Time, Voltage, Current, Charge, Power).
9. Tools for Columbic efficiency analysis, current interrupt, rest time, and multiple loop cycles.
10. Profile importation for Urban Life Cycle testing.
11. Analysis tools for batteries, fuel cells, corrosion, photovoltaics, supercapacitors, advanced electrochemistry etc.


